

DO Meter Quick Guide

Models	Parameters	Range	Accuracy	Cal. Points	GLP	Memory	Barometric / Salinity Compensation	Housing Rating	Communication	Page No.
Handheld Meters										
CyberScan PCD 650	pH/mV/Ion/Conductivity/TDS/Salinity/Resistivity/Dissolved Oxygen/°C/°F	0.00 to 90.00 mg/L or ppm	±0.2 mg/L	1 (mg/L)	Yes	500	Yes	IP67	LED RS232, IrDA	Pg 68
CyberScan PD 650	pH/mV/Ion/Dissolved Oxygen/°C/°F	0.00 to 90.00 mg/L or ppm	±0.2 mg/L	1 (mg/L)	Yes	500	Yes	IP67	LED RS232, IrDA	Pg 70
CyberScan CD 650	Conductivity/TDS/Salinity/Resistivity/Dissolved Oxygen/°C/°F	0.00 to 90.00 mg/L or ppm	±0.2 mg/L	1 (mg/L)	Yes	500	Yes	IP67	LED RS232, IrDA	Pg 71
CyberScan DO 600	Dissolved Oxygen/°C/°F	0 to 600 %, 0 to 90 mg/L	±2 %, ±0.2 mg/L	2 (0 %, 100 %) 1 (mg/L)	Yes	500	Yes	IP67	LED RS232, IrDA	Pg 47
CyberScan DO 300	Dissolved Oxygen/°C/°F	0 to 199.9 %, 0 to 19.99 mg/L	±1.5 % Full scale	2 (0 %, 100 %) 1 (mg/L)	Yes	50	Yes	IP67	–	Pg 48
CyberScan DO 110	Dissolved Oxygen/°C/°F	0 to 199.9 %, 0 to 19.99 mg/L	±1.5 % Full scale	2 (0 %, 100 %) 1 (mg/L)	–	100	Yes	IP54	RS232	Pg 49
EcoScan DO 6	Dissolved Oxygen/°C	0 to 199.9 %, 0 to 19.99 mg/L	±1.5 % Full scale	2 (0 %, 100 %) 1 (mg/L)	–	–	Yes	IP54	–	Pg 50
Bench Meters										
CyberScan PCD 6500	pH/°C/°F/mV/Ion/Conductivity/TDS/Salinity/Resistivity/Dissolved Oxygen/BOD/OUR/SOUR	0 to 600 %, 0 to 60 mg/L 0 to 1272 mbar	0.1 % 0.1 +1 LSD 0.1 mbar	2 (0 %, 100 %) 2 (0 mg/L, 8.26 mg/L)	Yes	1000 Per parameter	Yes	IP54	RS232, USB, IrDA	Pg 76
CyberScan DO 6000	Dissolved Oxygen/BOD/OUR/SOUR/°C/°F	0 to 600 %, 0 to 60 mg/L 0 to 1272 mbar	0.1 % 0.1 +1 LSD 0.1 mbar	2 (0 %, 100 %) 2 (0 mg/L, 8.26 mg/L)	Yes	1000 Per parameter	Yes	IP54	RS232, USB, IrDA	Pg 52
CyberScan DO 1500	Dissolved Oxygen/°C/°F	0 to 600 %, 0 to 60 mg/L	±0.5 % Full scale +1 LSD	2 (0 %, 100 %) 1 (mg/L)	–	100	Yes	IP54	RS232	Pg 53

